

AREA	SUBAREA	DEVICE	LEVEL OF DISINFECTION	FREQUENCY	WHO	DISINFECTING AGENT
		Spuds, Alger brush, forceps, lid tools etc	High		1 OD	High - listed below
		Tonometer/ pachymeter probe	Manufacturer recommendations		1 OD	
		Gonioscopy lens	Manufacturer recommendations		1 OD	
		Diagnostic lenses	Manufacturer recommendations		1 OD	
		Forehead/chin rests, breath shields	Low		1 OD or staff	Low- listed below
		Occluders, stereo glasses	Low		1 OD	Low- listed below
	<b>EXAM ROOM</b>	Diagnostic equip (OCT, VF, fundus camera)	Manufacturer recommendations		1 Staff	
		ICare Tonometer (disposable probe)	Handle - Low		1 Staff	Low- listed below
		Exam chair and unit (table, phoropter etc)	Low		1 OD	Low- listed below
		Trial Frame, trial lenses	Low		1 OD or staff	Low- listed below
		Other handhelds (rulers etc)	Low		1 OD	Low- listed below
		Sinks, Counters, Taps etc	Low		2 OD	Low- listed below
<b>PROFESSIONAL</b>						
		Frames on display	Low		1 Staff	Detergent and water
		Frame displays	Low		5 Staff	Low- listed below
		Frame warmer	Low		3 Staff	Low- listed below
	<b>LAB/DISPENSING</b>	Lab hand tools	Low		1 Staff	Low- listed below
		Plexiglass shield	Low		1 Staff	Low- listed below
		Optikam frame	Low		1 Staff	UV
		Pupillometer	Low		1 Staff	Low- listed below
		Sink, Counters, Taps etc	Low		2 Staff	Low- listed below
		Desk counters, staff chairs	Low		2 Staff	Low- listed below
		Computer keyboards/mouse, telephone	Low		2 Staff	Low- listed below
<b>ADMINISTRATIVE</b>		POS terminal	Low		1 Staff	Low- listed below
		Stapler, personal pens	Low		2 Staff	Low- listed below
		Pens used by patient	Low		1 Staff	Low- listed below
		Fax machine, scanner	Low		3 Staff	Low- listed below
		Waiting area	Low		2 Staff	Low- listed below
<b>GENERAL OFFICE</b>		Door handles	Low		1 Staff	Low- listed below
		Light switches	Low		2 Staff	Low- listed below
		Front desk counter, plexiglass	Low		2 Staff	Low- listed below
		Chairs - patient use	Low		1 Staff	Low- listed below
<b>SOP Legend:</b>	Frequency:	1 - after each direct patient contact	Disinfecting Agents:	Low: include	Alcohols (75-90% ethyl or isopropyl alcohol)	
		2 - mid-day and end-of day			QUAT (quaternary ammonium cation)	
		3 - end-of-day			Phenolics (ie Pine Sol, Lysol)	
		4 - weekly			CaviCide	
		5 - monthly		High: include	2% glutaraldehyde	
					0.2% peracetic acid	
					3% hydrogen peroxide	
					0.55% ortho-phthalaldehyde (OPA)	